CDF Operations Report

Andy Hocker, University of Rochester

23-JUN-2003

All Experimenters' Meeting

Stores summary

Date	Store	Duration (h)	Inst Lum (initial)	Int Lum (delivered)	Int Lum (live)	ε
Mo 6/16	2692	15.1	27.4e30	890.5	789.9	88.7%
Tu 6/17	2694	17.2	31.5e30	1077.9	837.9	77.7%
We 6/18	2702	5.2	29.8e30	463.4	421.6	90.9%
Th 6/19	2705	13.2	22.8e30	671.5	451.6	67.2%
Fr 6/20	2707	6.2	29.1e30	506.2	474.5	93.7%
Fr 6/20	2710	17.1	24.7e30	950.6	865.4	91.0%
Sa 6/21	2712	17.6	30.7e30	1119.3	1028.7	91.9%
Su 6/22	2715	20.2	39.3e30	1588	1447	91.1%
Total				7267.4	6316.6	86.9%

Down time

- Fan failed in PS for track trigger electronics crate Thursday evening
  - Tried to revive, then had to call in expert to swap out PS
  - Two hours
- Sacrificed beam time Tu and Th for "two SRC tests"
  - SVX/ISL/L00 read out every L1A, only SVX used in L2 trigger
  - Give ISL/L00 their own Silicon Readout Controller and only read out after L2A
  - 2nd SRC does not play well with TDCs --- only reproducible with beam --- enlisted this guy's help
  - Test results inconclusive, more offline head-scratching needed

Why 2 SRCs?



silicon readout times



resonance probability (n readouts in a row)

- Benefits of two SRCs:
  - Shorter L2 decision time (8 μs)
  - Less probability for readout resonances (=broken wire bonds) at high L1A rates

Sunday access

- Dead wedge in plug hadron calorimeter discovered in between-store calibration
  - Expert diagnosed blown fuse in collision hall
- Data missing a wedge deemed "bad" --- pbars not yet loaded --- request access
  - Silicon tags along, as usual



Post-access: Fully-functional calorimeter, two fixed Si ladders, and highest initial luminosity in 35 days

Summary

- When not testing, CDF had a week of smooth running
  - We like D0's "do not touch" policy
  - 6.3 pb^{-1} to tape, 87% efficiency
- Timing of blown calorimeter fuse was unfortunate
  - We appreciate being given the opportunity to take full advantage of TeV's "return to form"
  - All heckling graciously acknowledged
- Looking forward to more high luminosity running
  - Time has been invested in the past month to ensure we are ready to take data efficiently at high rate